

(2) is certified or meets the requirements for certification in a medical specialty in which he has become qualified by experience and training in the use of X-rays for diagnostic purposes, or (3) specializes in radiology and is recognized by the medical community as a specialist in radiology.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

**§ 486.104 Condition for coverage: Qualifications, orientation and health of technical personnel.**

Portable X-ray services are provided by qualified technologists.

(a) *Standard—qualifications of technologists.* All operators of the portable X-ray equipment meet the requirements of paragraphs (a)(1), (2), (3), or (4) of this section:

(1) Successful completion of a program of formal training in X-ray technology in a school approved by the Joint Review Committee on Education in Radiologic Technology (JRCERT), or have earned a bachelor's or associate degree in radiologic technology from an accredited college or university.

(2) For those whose training was completed prior to July 1, 1966, but on or after July 1, 1960: Successful completion of 24 full months of training and/or experience under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation experience in the 5 years prior to January 1, 1968.

(3) For those whose training was completed prior to July 1, 1960: Successful completion of 24 full months of training and/or experience of which at least 12 full months were under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation ex-

perience in the 5 years prior to January 1, 1968.

(4) For those whose training was completed prior to January 1, 1993, successful completion of a program of formal training in X-ray technology in a school approved by the Council on Education of the American Medical Association, or by the American Osteopathic Association is acceptable.

(b) *Standard—personnel orientation.* The supplier of portable X-ray services has an orientation program for personnel, based on a procedural manual which is: Available to all members of the staff, incorporates relevant portions of professionally recognized documents, and includes instruction in all of the following:

(1) Precautions to be followed to protect the patient from unnecessary exposure to radiation;

(2) Precautions to be followed to protect an individual supporting the patient during X-ray procedures from unnecessary exposure to radiation;

(3) Precautions to be followed to protect other individuals in the surrounding environment from exposure to radiation;

(4) Precautions to be followed to protect the operator of portable X-ray equipment from unnecessary exposure to radiation;

(5) Considerations in determining the area which will receive the primary beam;

(6) Determination of the time interval at which to check personnel radiation monitors;

(7) Use of the personnel radiation monitor in providing an additional check on safety of equipment;

(8) Proper use and maintenance of equipment;

(9) Proper maintenance of records;

(10) Technical problems which may arise and methods of solution;

(11) Protection against electrical hazards;

(12) Hazards of excessive exposure to radiation.

(c) *Standard: Employee records.* Records are maintained and include evidence that—

(1) Each employee is qualified for his or her position by means of training and experience; and

## § 486.106

(2) Employees receive adequate health supervision.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977, and amended at 53 FR 12015, Apr. 12, 1988; 60 FR 45086, Aug. 30, 1995; 73 FR 69942, Nov. 19, 2008]

### § 486.106 Condition for coverage: Referral for service and preservation of records.

All portable X-ray services performed for Medicare beneficiaries are ordered by a doctor of medicine or doctor of osteopathy and records are properly preserved.

(a) *Standard—referral by a physician.* Portable X-ray examinations are performed only on the order of a doctor of medicine or doctor of osteopathy licensed to practice in the State. The supplier's records show that:

(1) The X-ray test was ordered by a licensed doctor of medicine or doctor of osteopathy, and

(2) Such physician's written, signed order specifies the reason an X-ray test is required, the area of the body to be exposed, the number of radiographs to be obtained, and the views needed; it also includes a statement concerning the condition of the patient which indicates why portable X-ray services are necessary.

(b) *Standard—records of examinations performed.* The supplier makes for each patient a record of the date of the X-ray examination, the name of the patient, a description of the procedures ordered and performed, the referring physician, the operator(s) of the portable X-ray equipment who performed the examination, the physician to whom the radiograph was sent, and the date it was sent.

(c) *Standard—preservation of records.* Such reports are maintained for a period of at least 2 years, or for the period of time required by State law for such records (as distinguished from requirements as to the radiograph itself), whichever is longer.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

## 42 CFR Ch. IV (10–1–10 Edition)

### § 486.108 Condition for coverage: Safety standards.

X-ray examinations are conducted through the use of equipment which is free of unnecessary hazards for patients, personnel, and other persons in the immediate environment, and through operating procedures which provide minimum radiation exposure to patients, personnel, and other persons in the immediate environment.

(a) *Standard—tube housing and devices to restrict the useful beam.* The tube housing is of diagnostic type. Diaphragms, cones, or adjustable collimators capable of restricting the useful beam to the area of clinical interest are used and provide the same degree of protection as is required of the housing.

(b) *Standard—total filtration.* (1) The aluminum equivalent of the total filtration in the primary beam is not less than that shown in the following table except when contraindicated for a particular diagnostic procedure.

Operating kVp	Total filtration (inherent plus added)
Below 50 kVp .....	0.5 millimeters aluminum.
50–70 kVp .....	1.5 millimeters aluminum.
Above 70 kVp .....	2.5 millimeters aluminum.

(2) If the filter in the machine is not accessible for examination or the total filtration is unknown, it can be assumed that the requirements are met if the half-value layer is not less than that shown in the following table:

Operating kVp	Half-value layer
50 kVp .....	0.6 millimeters aluminum.
70 kVp .....	1.6 millimeters aluminum.
90 kVp .....	2.6 millimeters aluminum.
100 kVp .....	2.8 millimeters aluminum.
110 kVp .....	3.0 millimeters aluminum.
120 kVp .....	3.3 millimeters aluminum.

(c) *Standard—termination of exposure.* A device is provided to terminate the exposure after a preset time or exposure.

(d) *Standard—control panel.* The control panel provides a device (usually a milliammeter or a means for an audible signal to give positive indication of the production of X-rays whenever the X-ray tube is energized. The control panel includes appropriate indicators (labelled control settings and/or meters) which show the physical factors